



**INDIAN INSTITUTE OF SCIENCE EDUCATION AND
RESEARCH THIRUVANANTHAPURAM (IISERTVM)
(Govt. of India, Ministry of H.R.D)**

PH: 0471-2597454
FAX: 0471-2597427
EMAIL: PURCHASESTORES@IISERTVM.AC.IN

CET CAMPUS, KULATHUR, ENGINEERING COLLEGE.P.O
THIRUVANANTHAPURAM 695016,
KERALA, INDIA.

Date: 23rd March 2015

INVITATION TO TENDER(IND)

No: IISER/PUR/5506/14

Due Date: 07th April 2015 [4PM]

Date of Opening: 08th April 2015 [3PM]

Dear Sirs,

Sub: Supply of **Single Channel Pipettes.**

We require the following items immediately:

Sl No:	Item/ Description	Qty
1.	Single Channel Pipettes [Specifications as per Annexure I]	
	2.5µl	6 Nos
	20µl	6 Nos
	200µl	6 Nos
	1000µl	6 Nos
	100µl	3 Nos
	5000µl	1 No

Please quote your lowest rate and shortest delivery period as per the following terms. Your offer in sealed cover **SUPERSCRIBING TENDER NUMBER AND DUE DATE** shall reach us on or before the due date and time. Please follow "Instruction to Tenderers" attached.

- Payment** : Within 30 days after supply and installation/Net 30 days /LC. No advance payment will be made by IISERTVM.
- Delivery** : To be delivered at our stores (Free delivery). If import, mention Ex-works/FCA/CIP terms with clear breakup charges.
- Warranty** : For a minimum period of one year from the date of supply and installation
- Taxes & Duties** : Indicate taxes and duties. We are exempted for customs duty under 51/96 notification and Excise duty under 10/97 notifications.
- Discount** : Indicate, if any.
- Delivery Schedule** : Required Immediately. Indicate your schedule
- Validity of quote** : 60 days

Thanking You,

Yours Faithfully



K. BHASKARA RAO
Deputy Registrar
[Purchase & Stores]

P.S: CATALOGUE/LEAFLET FEATURING ALL TECHNICAL SPECS/INFORMATION OF THE PRODUCT QUOTED SHALL ACCOMPANY THE OFFER.

ANNEXURE I TO ENQUIRY
No. IISER/PUR/5506/14
DATED 23RD MARCH 2015

Specifications:

1. Volume ranges upto 2.5 μ l, 20 μ l, 100 μ l, 200 μ l, 1000 μ l and 5000 μ l
2. Volume should be adjustable in four digit range.
3. Should have user friendly piston system and tip ejector system.
4. Easily cleanable
5. Autoclavable
6. Should be resistant to UV exposure
7. Should have 3 year warranty
8. Should be able to provide, service and calibration periodically.
9. Should have provided the pipettes to National Institutes of repute (should provide the list).
10. Quote separately for pipette stand/individual pipette holders as accessories.


K. BHASKARA RAO
Deputy Registrar
[Purchase & Stores]