

Date: 16th December 2016

ADDENDUM -II TO IFT NO

No: IISER/PUR/8545/16

**Sub: Supply of Total Internal Reflection Fluorescence (TRIF)
Microscope.**

The above referred tender specifications is hereby revised as below:

7b. At least the following laser lines should be present:

- i) Diode/DPSS 488nm \geq 40mW
- ii) Diode/DPSS 561nm \geq 40mW

**Also please note that the Due date for submitting quotation is
extended to 12th January 2017 [4PM].**

Date of Opening Quotation is 13th January 2017 [3PM].

All other T&C of the IFT remains unchanged

Thanking You,



Yours Faithfully

Deputy Registrar (I/C)
Purchase & Stores

SUDIN B BABU
Deputy Registrar (I/C)
[Purchase & Stores]

Indian Institute of Science Education and Research Thiruvananthapuram
CET Campus, Engineering College P.O.
Thiruvananthapuram-695 016