



**INDIAN INSTITUTE OF SCIENCE EDUCATION AND
RESEARCH THIRUVANANTHAPURAM [IISERTVM]**

PH.-0471 2597454,
FAX: 0471-2597427
EMAIL: purchasestores@iisertvm.ac.in

CET CAMPUS, ENGINEERING COLLEGE. P.O
THIRUVANANTHAPURAM 695016,
KERALA, INDIA

Date: 02nd March 2016

ADDENDUM TO TENDER NO

No: IISER/PUR/7141/15

Sub: Supply of Rotary Evaporator.


The Invitation to Tender for Rotary Evaporator should include the Recirculating Chiller with following specifications also.

- Each chiller should be connected to one rotavapor only
- No of chiller required: 5
- Cooling capacity at 15°C – 550W or better
- Temperature range - -10°C to +25°C
- Frequency : 230V, 50/60 Hz
- Temperature Display – resolution 0.1 °C or better
- Refrigerant – R-134a, CFC free refrigerant.
- Temperature regulation accuracy - +/-1 °C or lower
- Tank volume – 3.0 L or more
- Flow Rate: 2.5 L/min

Other terms and conditions of our invitation for bids remain same.

Thanking You,

Yours Faithfully


K. Bhaskara Rao 2/3/16
Deputy Registrar
(Purchase & Stores)