



**INDIAN INSTITUTE OF SCIENCE EDUCATION AND
RESEARCH THIRUVANANTHAPURAM [IISERTVM]**

PH.-0471 2597454,
FAX: 0471-2597427
EMAIL: purchasestores@iisertvm.ac.in

CET CAMPUS, ENGINEERING COLLEGE. P.O
THIRUVANANTHAPURAM 695016,
KERALA, INDIA

Date: 04th December 2015

ADDENDUM TO TENDER NO

No: IISER/PUR/6599/15

Dear Sirs,


Sub: Weighing Balance.

The following specifications in the Tender may be read as follows instead of existing one:

Weighing Balance with the accuracy to measure 1 mg i.e., 3 digit decimal in the scale of 'gm'. Capacity- 300gm to 500gm

Thanking You,

Yours Faithfully
Sd/-
(Competent Authority)
IISERTVM

Approved for Issue

**AjithPrabha
MSA [P&S]**
4/12/15