



**INDIAN INSTITUTE OF SCIENCE EDUCATION AND  
RESEARCH THIRUVANANTHAPURAM [IISERTVM]**

PH.-0471 2597454,  
FAX: 0471-2597427

EMAIL: [purchasestores@iisertvm.ac.in](mailto:purchasestores@iisertvm.ac.in)

CET CAMPUS, ENGINEERING COLLEGE. P.O  
THIRUVANANTHAPURAM 695016,  
KERALA, INDIA

Date: 27<sup>th</sup> April 2017

**ADDENDUM TO TENDER NO**

**No: IISER/PUR/0026/AT/SC/17-18**

**SUB: Supply of Electrochemical Work Station**

Please **read the Specification No. 6) At least 400 mA to 100 nA or lower (and at least one channel should have the range of pA to 400 mA)** instead of At least 400 mA to 100 nA or lower (and at least one channel should have the range of 1pA to 400 mA) **& Specification No. 8) 800 pA or better (800 fA or better for pA range)** instead of 800 pA or better (800 fA or better for 1pA range)

Rest of the things remain unchanged.

Thanking You,

Yours Faithfully

Deputy Registrar  
In-Charge  
[Purchase & Stores]