



**INDIAN INSTITUTE OF SCIENCE EDUCATION AND
RESEARCH THIRUVANANTHAPURAM [IISERTVM]**

PH.-0471 2778019,
EMAIL: purchasestores@iisertvm.ac.in

MARUTHAMALA.P.O VITHURA. P.O
THIRUVANANTHAPURAM 695551,
KERALA, INDIA
GST No.32AAAJI0299R1ZS

Date: 13/12/2018

**CORRIGENDUM-I TO TENDER No
IISER/PUR/0222/SST/SB/18-19**

**Sub: Supply of Upright Biomedical freezer.
Ref: Tender ID -2018-IISRT-390177_1**

As the bids received are not technically qualified the above referred tender is hereby retendered with revised specification as per annexure 1(enclosed).

Please note the new dates as below:

**Due Date: 26/12/2018 [3PM]
Date of Opening: 27/12/2018 [3.30PM]**


All other Terms and Conditions are remains same.

Thanking You,

Yours Faithfully

Deputy Registrar
Purchase & Stores

Technical Specifications for Upright Biomedical Freezer:	
1	Should be a frost free upright freezer for storing biological samples at temperature up to -30°C controllable from -20 to -30°C
2	Should have a capacity of at least 450 litres or more.
3	Should have two independent compartments with independent doors (double-door model)
4	Should have microprocessor-based controller with LED display for inside temperature
5	Should have front mounted LED display and digital input control panel on the exterior
6	Should have manual defrost model
7	Should come with at least 5-day temperature recorder
8	Should have CFC free refrigerant
9	The exterior measurement should not exceed 190cm (width) x 100cm (depth)
10	Noise level should be less than 45db
11	Should come with painted steel exterior and thermoplastic interiors or painted steel interiors
12	Should have an insulation of at least 65mm thickness
13	Should come with at least 6 stainless steel racks that fit 2" boxes for one compartment
14	Should come with at 6 drawers for storage in the other compartment
15	Should have at least one access port
16	Electrical requirement should be 230V and 50Hz
17	Should have alarm: <ul style="list-style-type: none"> - high /low temperature - power failure - door left open - low battery
18	Should come with lockable caster wheels for easy movement along with permanent leg
19	The bidder should have prior successful installations in at least 25 Institutes of National Importance of the quoted equipment (or upright biomedical freezers similar to the quoted model, but identical make) and should provide the list
20	The equipment should be CE and /or UL listed
21	The net weight should be in range of 100-200 Kg
22	Should come with a warranty of at least 3 years
23	Supplier should have Indian service provider



Deputy Registrar
[Purchase & Stores]