

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

THIRUVANANTHAPURAM [IISERTVM]

PH.-0471 2597454,

FAX: 0471-2597427

EMAIL: purchasestores@iisertvm.ac.in

CET CAMPUS, ENGINEERING COLLEGE.P.O

THIRUVANANTHAPURAM 695016,

KERALA, INDIA

IISER/PUR/4506/14

3rd September 2014

CORRIGENDUM II

Point No. 1.3 of Annexure I of our Corrigendum dated 22nd August 2014 has been amended to the following:

Internal side wall panels shall be fabricated using 6 mm thick liner of Trespa® TopLab^{PLUS}® / Trespa Athlon Make. The work top shall be made of 20 mm thick material of Trespa® TopLab^{PLUS}® make and shall be dish type or have skirting of approx. 10 mm. Worktop shall be placed at a height of 900 mm from the floor level.

Rest remains unchanged.

**Deputy Registrar [F&A]
Addl. Charge [P&S]**