

Date: 27<sup>th</sup> January 2017

**ADDENDUM TO TENDER NO:**

**IISER/PUR/9157/SK/SB/16**

**Sub: Supply of Table Top Non-Refrigerated Centrifuge**

The above referred tender specifications is hereby revised as below:

- Should be a compact, table top centrifuge
- Centrifuge should have maximum Speed at least 20000xg
- Should come with a fixed angle rotor with more than 20 places for 1.5/2ml microcentrifuge tubes
- An additional rotor to adjust spin columns for molecular biology applications should also be provided.
- The vendor should also supply adaptors for 0.5 ml microcentrifuge tubes for all the available places in the rotor
- Maximum power requirement should be less than 500W
- Noise level should be less than 65 db
- Short spin key with selectable rotational speed
- Should come with spill free lid with mechanical lock
- Should have automatic imbalance detection and other operator safety features
- Centrifuge timer should start after the set RPM is reached
- Clear electronic display for all parameters
- Should provide three years warranty
- Three years AMC, post warranty period should be quoted separately
- Should provide users list, research institutes
- Should provide printed brochures with detailed specifications
- The vendor should dismantle and reinstall the whole system from our present transit campus to permanent campus at Vithura with no additional cost
- Should be equipped with energy efficient features. Should have auto switch off mode when not in use for reduced energy consumption and extended compressor life
- Compliance sheet for all specification should be attached. Wherever relevant, the appropriate section/page number of technical brochure should be cited.


**Also please note that the Due date for submitting quotation is extended to 21<sup>st</sup> February 2017 [4PM].**

**Date of Opening Quotation is 22<sup>nd</sup> February 2017 [3PM].**

All other Terms & Conditions remains unchanged.

Thanking You,

Yours Faithfully



Asst. Registrar  
Gen Admn