

Date: 22 Nov 2023

**ADDENDUM-II TO TENDER NUMBER
IISER/PUR/1146/MT-P/SP/23-24**

**Sub: Supply and Installation of Fume Hood and Laminar Flow Bench: reg
Ref: CPPP Tender ID 2023_IISRT_780991_1**

1. The updated technical specifications are as follows:-

Technical specifications

A. Fume hood: Floor standing model

- i. The working space area: 90X60X60 cm
- ii. The work space should be Acid/alkali resistant
- iii. Construction: Exterior: Steel with powder coating or wooder with laminated mica finish. Interior : laminate board or other good quality materials with good acid and alkali resistance. Work Table: Chemical resistant resin or other comparable material with good acid and alkali resistance.
- iv. The slash/door should be vertically sliding
- v. Should be equipped with a proper exhaust system to handle the acid/alkali fumes
- vi. Work space should have standard power sockets, adequate lighting.
- vii. Workspace should have water tap and a small sink
- viii. Should operate with standard 220/230V.
- ix. Bottom cupboard with doors: 02
- x. Should provide appropriate floor-stand to mount the fume hood

B. Laminar flow bench

- i. Cleanliness: Class 5-or better
- ii. Pre-filter down to approx. 10-microns and HEPA Filter down to 0.5 microns or lower
- iii. Configuration: Vertical with standing slash.
- iv. Work space: 90 x 60 x 60 cm, Stainless steel top
- v. Construction: Steel powder coated outside
- vi. Bottom cupboard with doors: 02
- vii. Illumination: LED/ Tube light
- viii. Standard 220/230 V, 6A Power outlets inside to be provided
- ix. Compatible floor stand with castors.


2 The due date is hereby extended as follows:-

Due Date: 05 Dec 2023 1500 Hrs
Date of opening 06 Dec 2023 1530 Hrs

3. All other Terms and Conditions remain the same. Bidders may quote accordingly

Thanking You,




21/11/2023
Yours Faithfully
Deputy Registrar (P&S)