



**INDIAN INSTITUTE OF SCIENCE EDUCATION AND
RESEARCH THIRUVANANTHAPURAM [IISERTVM]**

PH.-0471 2778019,
EMAIL: purchasestores@iisertvm.ac.in

MARUTHAMALA.P.O VITHURA. P.O
THIRUVANANTHAPURAM 695551,
KERALA, INDIA
GST No.32AAAJI0299R1ZS

Date: 03/12/2019

CORRIGENDUM TO TENDER NO

No: IISER/PUR/0976/RM/SB/19-20

Sub: Supply of Gel Documentation System.
Ref: Tender Enquiry No. 2019_IISRT_514587_1

The above-referred tender enquiry has been **RETENDERED** with amended technical specification as mentioned in Annexure 1

2. Accordingly, the due date for submitting and date of opening of bids has been revised as given below:


Due date for submission : 17.12.2019 (3.00 PM)
Date of Opening : 18.12.2019 (3.30 PM)

3. Bidders are requested to upload their technical specifications & compliance sheet taking into consideration the above amendment.

4. **All other terms and conditions remains same.**

Thanking You,

Yours Faithfully


Deputy Registrar
Purchase & Stores

Annexure I to IISER/PUR/0976/RM/SB/19-20

Technical Specifications	
1	Should be a stand-alone Gel documentation system for fluorescence and white light imaging of DNA, RNA and protein.
2	The system should have in-built processor with multi touch capability with 12" touch screen
3	Should have a Scientific grade CCD Camera of 16 bit 65536 grey scale or better with at least 5 megapixel resolution.
4	Should have motorized zoom lens with at least f/1.2
5	Should have auto exposure and auto illumination features
6	System should have an automatic control of the camera, lens and lighting.
7	Image acquisition should be done with one click.
8	Should have in built software for image editing and image analysis
9	Data output format should be at least in TIFF and JPEG.
10	Should have network connectivity.
11	The vendor should provide a Thermal printer with the high quality (ultra sound imaging quality) direct thermal printing technology and media (20 roll or equivalent for each system) along with the system.
12	Should provide 5 years warranty from the date of installation for complete system including illumination sources.
13	Compliance sheet for all specification should be attached. Where ever relevant, the appropriate section/page number of technical brochure should be cited.

[Handwritten signature]
03/12/19